

ORE RESEARCH & EXPLORATION P/L ABN 28 006 859 856 37A Hosie Street · Bayswater North · VIC 3153 · AUSTRALIA ▲ 61 3 9729 0333 <sup>F</sup> 61 3 9729 8338 i) info@ore.com.au ⊕www.ore.com.au

#### **CERTIFICATE OF ANALYSIS FOR**

# HIGH SULPHIDATION EPITHERMAL Ag-Cu-Au ORE **CERTIFIED REFERENCE MATERIAL OREAS 603**

	Jes, 5D s, 9		ence and TC				
Constituent	Certified	1SD	95% Confid	lence Limits	95% Tolera	ance Limits	
	Value Low		Low	High	Low	High	
Fire Assay							
Ag, Silver (ppm)	284	15.9	275	294	278	291	
Au, Gold (ppm)	5.18	0.151	5.12	5.23	5.14*	5.21*	
4-Acid Digestion							
Ag, Silver (ppm)	298	8.1	294	301	292	303	
Al, Aluminium (wt.%)	3.98	0.156	3.90	4.06	3.89	4.08	
As, Arsenic (ppm)	1801	119.1	1743	1858	1760	1842	
Be, Beryllium (ppm)	0.71	0.08	0.67	0.75	0.66	0.76	
Bi, Bismuth (ppm)	149	13.2	143	155	144	155	
Ca, Calcium (wt.%)	0.318	0.015	0.312	0.324	0.307	0.329	
Cd, Cadmium (ppm)	54	3.3	52	55	52	55	
Ce, Cerium (ppm)	25.7	3.7	22.2	29.2	24.0	27.5	
Co, Cobalt (ppm)	15.3	0.96	14.9	15.7	14.8	15.8	
Cr, Chromium (ppm)	30.2	5.3	27.8	32.7	27.4	33.0	
Cs, Cesium (ppm)	1.66	0.071	1.61	1.72	1.56	1.76	
Cu, Copper (wt.%)	1.00	0.034	0.99	1.02	0.99	1.02	
Dy, Dysprosium (ppm)	1.21	0.072	1.15	1.27	1.13	1.29	
Er, Erbium (ppm)	0.59	0.024	0.58	0.61	0.54	0.65	
Eu, Europium (ppm)	< 1	IND	IND	IND	IND	IND	
Fe, Iron (wt.%)	2.92	0.135	2.86	2.98	2.85	2.99	
Ga, Gallium (ppm)	22.2	1.96	21.1	23.3	21.6	22.9	
Gd, Gadolinium (ppm)	1.77	0.107	1.69	1.85	1.54	2.00	
Hf, Hafnium (ppm)	2.53	0.167	2.42	2.63	2.37	2.68	

\*Gold Tolerance Limits for typical 30g fire assay charge weight determined from 20 x 1g NAA results and the Sampling Constant (Ingamells & Switzer, 1973).

Please note: intervals may appear asymmetric due to rounding.



Table 1 continued.								
Constituent	Certified	190	95% Confid	lence Limits	95% Tolerance Limits			
Constituent	Value	100	Low	High	Low	High		
4-Acid Digestion continued								
Ho, Holmium (ppm)	0.21	0.020	0.19	0.23	0.17	0.25		
In, Indium (ppm)	11.2	0.96	10.7	11.7	10.8	11.6		
K, Potassium (wt.%)	0.623	0.019	0.614	0.632	0.606	0.639		
La, Lanthanum (ppm)	11.9	2.3	10.5	13.3	11.1	12.7		
Li, Lithium (ppm)	18.9	1.30	18.2	19.7	18.3	19.6		
Lu, Lutetium (ppm)	0.099	0.012	0.088	0.109	IND	IND		
Mg, Magnesium (ppm)	828	45.4	809	848	806	851		
Mn, Manganese (ppm)	133	10.8	127	138	129	136		
Mo, Molybdenum (ppm)	6.05	0.516	5.83	6.27	5.73	6.38		
Na, Sodium (wt.%)	0.428	0.015	0.421	0.436	0.414	0.442		
Nb, Niobium (ppm)	7.09	0.74	6.55	7.64	6.81	7.38		
Nd, Neodymium (ppm)	11.4	1.1	10.1	12.7	10.8	11.9		
Ni, Nickel (ppm)	112	4.7	110	114	108	116		
P, Phosphorus (ppm)	534	31.3	523	545	510	558		
Pb, Lead (ppm)	1908	124.8	1852	1964	1851	1965		
Pr, Praseodymium (ppm)	3.23	0.190	3.01	3.44	3.04	3.41		
Rb, Rubidium (ppm)	23.6	1.91	22.3	25.0	22.7	24.5		
S, Sulphur (wt.%)	3.71	0.161	3.62	3.79	3.60	3.82		
Sb, Antimony (ppm)	205	17.7	196	214	197	213		
Sc, Scandium (ppm)	4.04	0.50	3.81	4.27	3.73	4.35		
Se, Selenium (ppm)	60	9	55	66	57	63		
Sm, Samarium (ppm)	2.14	0.134	2.00	2.27	2.00	2.27		
Sn, Tin (ppm)	12.9	0.98	12.3	13.5	12.1	13.6		
Sr, Strontium (ppm)	459	77	422	497	441	478		
Ta, Tantalum (ppm)	< 1	IND	IND	IND	IND	IND		
Tb, Terbium (ppm)	0.23	0.023	0.21	0.25	0.21	0.25		
Te, Tellurium (ppm)	56	7	49	63	54	59		
Th, Thorium (ppm)	5.90	0.85	5.28	6.52	5.54	6.27		
Ti, Titanium (wt.%)	0.191	0.018	0.182	0.200	0.185	0.197		
TI, Thallium (ppm)	4.18	0.174	4.07	4.29	4.02	4.35		
U, Uranium (ppm)	2.71	0.144	2.60	2.81	2.59	2.82		
V, Vanadium (ppm)	32.0	2.22	31.0	33.0	31.0	33.0		
W, Tungsten (ppm)	14.0	2.6	12.6	15.4	13.3	14.7		
Y, Yttrium (ppm)	5.64	0.431	5.43	5.85	5.42	5.86		
Yb, Ytterbium (ppm)	0.61	0.09	0.53	0.69	0.57	0.65		
Zn, Zinc (wt.%)	0.920	0.031	0.906	0.934	0.906	0.934		
Zr, Zirconium (ppm)	78	6.0	75	81	75	81		
Aqua Regia Digestion	T							
Ag, Silver (ppm)	293	12.9	286	300	289	297		
AI, Aluminium (wt.%)	0.580	0.092	0.537	0.624	0.565	0.595		
As, Arsenic (ppm)	1766	91.2	1722	1810	1732	1801		

Please note: intervals may appear asymmetric due to rounding.



Table 1 continued.								
Constituent	Certified	190	95% Confid	dence Limits	95% Tolerance Limits			
Constituent	Value	150	Low	High	Low	High		
Aqua Regia Digestion continu	led							
Au, Gold (ppm)	5.08	0.179	4.96	5.19	$5.03^{\dagger}$	5.12 <sup>†</sup>		
B, Boron (ppm)	< 10	IND	IND	IND	IND	IND		
Be, Beryllium (ppm)	< 0.5	IND	IND	IND	IND	IND		
Bi, Bismuth (ppm)	151	13.9	144	158	147	154		
Ca, Calcium (wt.%)	0.237	0.017	0.229	0.245	0.227	0.247		
Cd, Cadmium (ppm)	55	3.6	53	57	54	56		
Ce, Cerium (ppm)	13.0	2.4	11.2	14.7	12.4	13.5		
Co, Cobalt (ppm)	15.2	0.79	14.8	15.5	14.7	15.7		
Cr, Chromium (ppm)	29.2	2.47	28.2	30.2	26.9	31.6		
Cs, Cesium (ppm)	0.74	0.12	0.65	0.83	0.70	0.77		
Cu, Copper (wt.%)	1.01	0.026	1.00	1.02	0.99	1.03		
Dy, Dysprosium (ppm)	0.56	0.06	0.49	0.63	0.53	0.58		
Er, Erbium (ppm)	0.22	0.03	0.19	0.25	IND	IND		
Eu, Europium (ppm)	0.25	0.05	0.18	0.32	0.23	0.27		
Fe, Iron (wt.%)	2.85	0.155	2.78	2.92	2.79	2.91		
Ga, Gallium (ppm)	6.58	0.84	6.08	7.09	6.32	6.85		
Gd, Gadolinium (ppm)	0.86	0.12	0.74	0.98	0.81	0.92		
Hf, Hafnium (ppm)	0.46	0.07	0.40	0.52	0.43	0.49		
Hg, Mercury (ppm)	< 4	IND	IND	IND	IND	IND		
Ho, Holmium (ppm)	0.096	0.013	0.082	0.111	IND	IND		
In, Indium (ppm)	11.7	1.15	10.9	12.5	11.3	12.0		
K, Potassium (wt.%)	0.093	0.015	0.086	0.100	0.091	0.095		
La, Lanthanum (ppm)	< 10	IND	IND	IND	IND	IND		
Li, Lithium (ppm)	4.62	0.82	4.10	5.14	4.47	4.78		
Lu, Lutetium (ppb)	26	5	22	31	IND	IND		
Mg, Magnesium (ppm)	517	61	490	543	IND	IND		
Mn, Manganese (ppm)	130	6.5	127	133	126	133		
Mo, Molybdenum (ppm)	6.10	0.564	5.82	6.39	5.81	6.40		
Na, Sodium (ppm)	275	58	247	302	IND	IND		
Nb, Niobium (ppm)	< 0.5	IND	IND	IND	IND	IND		
Nd, Neodymium (ppm)	5.90	0.71	5.11	6.69	5.66	6.14		
Ni, Nickel (ppm)	113	3.2	111	114	110	116		
P, Phosphorus (ppm)	172	27	159	186	162	183		
Pb, Lead (ppm)	1691	36.3	1674	1707	1666	1715		
Pr, Praseodymium (ppm)	1.57	0.25	1.28	1.85	1.47	1.66		
Rb, Rubidium (ppm)	4.89	0.88	4.20	5.59	4.65	5.14		
S, Sulphur (wt.%)	3.46	0.185	3.36	3.56	3.38	3.53		
Sb, Antimony (ppm)	160	24	148	172	157	164		
Sc, Scandium (ppm)	1.06	0.16	0.98	1.15	1.03	1.10		
Se, Selenium (ppm)	59	8	54	64	57	61		
Sm, Samarium (ppm)	1.12	0.13	0.98	1.25	1.06	1.18		

<sup>†</sup>Gold Tolerance Limits for typical 25g aqua regia sample weight determined as above. Please note: intervals may appear asymmetric due to rounding.



		Table 1 c	ontinued.			
Constituent	Certified	160	95% Confid	dence Limits	95% Tolera	ance Limits
Constituent	Value	130	Low	High	Low	High
Aqua Regia Digestion continu	ed					
Sn, Tin (ppm)	12.0	1.11	11.3	12.7	11.7	12.3
Sr, Strontium (ppm)	46.4	6.1	43.0	49.8	44.7	48.0
Tb, Terbium (ppm)	0.11	0.02	0.10	0.13	0.10	0.12
Te, Tellurium (ppm)	57	2.8	55	60	56	59
Th, Thorium (ppm)	2.30	0.29	2.05	2.55	2.21	2.39
Ti, Titanium (ppm)	98	13	87	108	IND	IND
TI, Thallium (ppm)	4.07	0.60	3.62	4.52	3.89	4.25
U, Uranium (ppm)	0.89	0.10	0.81	0.96	0.84	0.93
V, Vanadium (ppm)	10.0	1.1	9.5	10.6	9.4	10.7
W, Tungsten (ppm)	3.88	0.78	3.20	4.57	3.66	4.11
Y, Yttrium (ppm)	2.28	0.35	2.04	2.52	2.21	2.35
Yb, Ytterbium (ppm)	0.17	0.02	0.16	0.19	IND	IND
Zn, Zinc (wt.%)	0.911	0.034	0.895	0.926	0.895	0.927
Zr, Zirconium (ppm)	16.0	1.6	15.0	17.0	15.5	16.6
Infrared Combustion						
S, Sulphur (wt.%)	3.97	0.161	3.88	4.06	3.91	4.03
Disease mater intermedia many anno		2. 1				

Please note: intervals may appear asymmetric due to rounding.

#### Table 2. Indicative Values for OREAS 603.

Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value
Pb Fire Assay		I		•	•			I
Pd	ppb	754	Pt	ppb	144			
Borate Fusion XRF								
Al <sub>2</sub> O <sub>3</sub>	wt.%	7.80	Fe <sub>2</sub> O <sub>3</sub>	wt.%	4.19	Pb	ppm	2015
As	ppm	1780	K <sub>2</sub> O	wt.%	0.773	SiO <sub>2</sub>	wt.%	75.29
Ва	ppm	11400	MgO	wt.%	0.180	Sn	ppm	12.5
CaO	wt.%	0.455	MnO	wt.%	0.020	SO <sub>3</sub>	wt.%	9.50
Со	ppm	25.0	Na <sub>2</sub> O	wt.%	0.550	TiO <sub>2</sub>	wt.%	0.335
Cr	ppm	35.0	Ni	ppm	95	U	ppm	< 10
Cu	ppm	10100	$P_2O_5$	wt.%	0.129	Zn	ppm	8800
Thermogravimetry								
LOI <sup>1000</sup>	wt.%	5.42						
Laser Ablation ICP-MS								
Ag	ppm	282	Ho	ppm	0.23	Sn	ppm	13.0
As	ppm	1715	In	ppm	10.1	Sr	ppm	571
Ва	ppm	11550	La	ppm	18.1	Та	ppm	0.71
Be	ppm	0.60	Lu	ppm	0.095	Tb	ppm	0.20
Bi	ppm	145	Mn	wt.%	0.014	Те	ppm	60
Cd	ppm	53	Мо	ppm	6.30	Th	ppm	6.98
Ce	ppm	30.7	Nb	ppm	7.79	Ti	wt.%	0.198
Со	ppm	16.5	Nd	ppm	11.5	TI	ppm	4.20
Cr	ppm	39.0	Ni	ppm	110	Tm	ppm	0.10
Cs	ppm	1.51	Pb	ppm	1945	U	ppm	2.73
Cu	ppm	9830	Pr	ppm	3.58	V	ppm	32.4



			Table 2 cont	inued						
Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value		
Laser Ablation ICP-MS	Laser Ablation ICP-MS continued									
Dy	ppm	1.20	Rb	ppm	23.3	W	ppm	14.8		
Er	ppm	0.69	Re	ppm	0.013	Y	ppm	6.28		
Eu	ppm	0.31	Sb	ppm	184	Yb	ppm	0.68		
Ga	ppm	24.1	Sc	ppm	4.35	Zn	ppm	8530		
Gd	ppm	1.66	Se	ppm	< 5	Zr	ppm	122		
Hf	ppm	3.64	Sm	ppm	2.04					
4-Acid Digestion										
В	ppm	< 20	Ge	ppm	5.05	Re	ppb	4		
Ва	ppm	2504	Hg	ppm	< 1	Tm	ppm	0.088		
Aqua Regia Digestion										
Ва	ppm	686	Pt	ppb	132	Та	ppm	< 0.01		
Ge	ppm	1.35	Re	ppb	1	Tm	ppm	0.028		
Pd	ppb	734	Ru	ppb	< 2					
Infrared Combustion	Infrared Combustion									
C	wt.%	0.068								

Note: the number of significant figures reported is not a reflection of the level of certainty of stated values. They are instead an artefact of ORE's in-house CRM-specific LIMS.

# INTRODUCTION

OREAS reference materials are intended to provide a low cost method of evaluating and improving the quality of analysis of geological samples. To the geologist they provide a means of implementing quality control in analytical data sets generated in exploration from the grass roots level through to prospect evaluation, and in grade control at mining operations. To the analyst they provide an effective means of calibrating analytical equipment, assessing new techniques and routinely monitoring in-house procedures.

# SOURCE MATERIALS

OREAS 603 was prepared from gold-silver-copper bearing ore from Evolution Mining's Mount Carlton Operation in Queensland, Australia and blended with argillic rhyodacite waste rock to achieve the desired grades. The mineralisation assemblage consists of pyrite, enargite/tennantite, tetrahedrite, digenite, covellite, sphalerite, galena, alunite, dickite, kaolinite and vuggy silica, hosted in advanced argillic altered rhyodacite containing sulphur-salts. OREAS 603 is one of a suite of six CRMs ranging in grades from 24ppm Ag, 0.2 ppm Au and 0.05% Cu to 980ppm Ag, 1.7ppm Au and 5.0% Cu.

# **COMMINUTION AND HOMOGENISATION PROCEDURES**

The material constituting OREAS 603 was prepared in the following manner:

- drying to constant mass at 105°C;
- crushing and milling of the barren material to 95% minus 75 microns;
- crushing and milling of the ore material to 100% minus 30 microns;
- blending in appropriate proportions to achieve the desired grades;



• packaging in 60g and 10g units sealed under nitrogen in laminated foil pouches.

# ANALYTICAL PROGRAM

Twenty eight commercial analytical laboratories participated in the program to certify the 117 elements reported in Table 1. The following methods were employed:

- Silver via 30-40g fire assay with gravimetric (11 labs) or ICP-OES (1 lab) finish;
- Gold via 20-40g\* fire assay with AAS (20 labs), ICP-OES (4 labs) or gravimetric (3 labs) finish;
- Instrumental neutron activation analysis for Au on 1g subsamples to confirm homogeneity (1 laboratory);
- Gold via 15-40g\* aqua regia digestion with ICP-MS (7 labs) or AAS (5 labs) finish;
- 4-Acid digestion for full elemental suite ICP-OES and ICP-MS (up to 21 laboratories depending on the element).
- Aqua regia digestion (see note below) for full elemental suite ICP-OES and ICP-MS (up to 22 laboratories depending on the element).
- Sulphur via Infrared Combustion Analysis (16 labs).

\*The certified values (and 95% Confidence Interval and SD) for Au are also applicable to 50g charge weights.

It is important to note that in the analytical industry there is no standardisation of the aqua regia digestion process. Aqua regia is a partial empirical digest and differences in recoveries for various analytes are commonplace. These are caused by variations in the digest conditions which can include the ratio of nitric to hydrochloric acids, acid strength, temperatures, leach times and secondary digestions. Recoveries for sulphide-hosted base metal sulphides approach total values, however, other analytes, in particular the lithophile elements, show greater sensitivity to method parameters. This can result in lack of consensus in an inter-laboratory certification program for these elements. The approach applied here is to report certified values in those instances where reasonable agreement exists amongst a majority of participating laboratories. The results of specific laboratories may differ significantly from the certified values, but will, nonetheless, be valid and reproducible in the context of the specifics of the aqua regia method in use. Users of this reference material should, therefore, be mindful of this limitation when applying the certified values in a quality control program.

For the round robin program twenty 1kg test units were taken at predetermined intervals during the bagging stage, immediately following final blending, and are considered representative of the entire batch. The six samples received by each laboratory were obtained by taking two 110g scoop splits from each of three separate 1kg test units. This format enabled nested ANOVA treatment of the results to evaluate homogeneity, i.e. to ascertain whether between-unit variance is greater than within-unit variance. Table 1 presents the certified values together with their associated 1SD's, 95% confidence and tolerance limits and Table 2 shows 89 indicative values for major and trace element composition. Gold homogeneity has been evaluated and confirmed by instrumental neutron activation analysis (INAA) on twenty ~1g sample portions (see Table 3 below) and by a nested ANOVA program for both fire assay and aqua regia digestion (see 'nested ANOVA' section). Table 4 provides performance gate intervals for the certified values based on their pooled 1SD's. Tabulated results of all elements (including Au INAA analyses) together with uncorrected means, medians, standard deviations, relative standard deviations and percent deviation of lab means from the corrected mean of



means (PDM<sup>3</sup>) are presented in the detailed certification data for this CRM (**OREAS 603 DataPack.xlsx**).

Replicate	NAA
No	1g
1	4.95
2	4.96
3	4.88
4	4.95
5	4.81
6	4.89
7	4.81
8	4.89
9	4.80
10	4.94
11	4.84
12	4.98
13	4.87
14	4.96
15	4.89
16	4.85
17	4.78
18	4.96
19	4.89
20	4.95
Mean	4.89
Median	4.89
Std Dev.	0.062
Rel.Std.Dev.	1.27%
PDM <sup>3</sup>	-5.47%

 Table 3. Neutron Activation Analysis of Au (ppm) on 20 x 1g subsamples.

## STATISTICAL ANALYSIS

**Certified Values, Confidence Limits, Standard Deviations and Tolerance Limits** (Table 1) have been determined for each analyte following removal of individual, laboratory dataset (batch) and 3SD outliers (single iteration). For individual outliers within a laboratory batch the z-score test is used in combination with a second method that determines the per cent deviation of the individual value from the batch median. Outliers in general are selected on the basis of z-scores > 2.5 and with per cent deviations (i) > 3 and (ii) more than three times the average absolute per cent deviation for the batch. In certain instances statistician's prerogative has been employed in discriminating outliers. Each laboratory data set mean is tested for outlying status based on z-score discrimination and rejected if > 2.5. After individual and laboratory data set (batch) outliers have been eliminated a non-iterative 3 standard deviation filter is applied, with those values lying outside this window also relegated to outlying status.

**Certified Values** are the means of accepted laboratory means after outlier filtering. The INAA data (see Table 3) is omitted from determination of the certified value for Au and is



used solely for the calculation of Tolerance Limits and homogeneity evaluation of OREAS 600.

**95% Confidence Limits** are inversely proportional to the number of participating laboratories and inter-laboratory agreement. It is a measure of the reliability of the certified value. A 95% confidence interval indicates a 95% probability that the true value of the analyte under consideration lies between the upper and lower limits. *95% Confidence Limits should not be used as control limits for laboratory performance.* 

**Indicative (uncertified) values** (Table 2) are provided for the major and trace elements determined by borate fusion XRF ( $AI_2O_3$  to Zn) and laser ablation with ICP-MS (Ag to Zr) and are the means of duplicate assays from Bureau Veritas, Perth. Additional indicative values by other analytical methods are present where the number of laboratories reporting a particular analyte is insufficient (< 5) to support certification or where inter-laboratory consensus is poor.

**Standard Deviation** values (1SDs) are reported in Table 1 and provide an indication of a level of performance that might reasonably be expected from a laboratory being monitored by this CRM in a QA/QC program. The SD values include all sources of measurement uncertainty: between-lab variance, within-run variance (precision errors) and CRM variability. For an effective CRM the contribution of the latter should be negligible in comparison to measurement errors. OREAS reference materials have a level of homogeneity such that the observed variance from repeated analysis has its origin almost exclusively in the analytical process rather than the reference material itself.

The SD for each analyte's certified value is calculated from the same filtered data set used to determine the certified value, i.e. after removal of any individual, lab dataset (batch) and 3SD outliers (single iteration). These outliers can only be removed after the absolute homogeneity of the CRM has been independently established, i.e. the outliers must be confidently deemed to be analytical rather than arising from inhomogeneity of the CRM. The standard deviation is then calculated for each analyte from the pooled accepted analyses generated from the certification program.

In the application of SD's in monitoring performance it is important to note that not all laboratories function at the same level of proficiency and that different methods in use at a particular laboratory have differing levels of precision. Each laboratory has its own inherent SD (for a specific concentration level and analyte-method pair) based on the analytical process and this SD is not directly related to the round robin program.

The majority of data generated in the round robin program was produced by a selection of world class laboratories. The SD's thus generated are more constrained than those that would be produced across a randomly selected group of laboratories. To produce more generally achievable SD's the 'pooled' SD's provided in this report include inter-lab bias. This 'one size fits all' approach may require revision at the discretion of the QC manager concerned following careful scrutiny of QC control charts.

Table 4 shows **Performance Gates** calculated for two and three standard deviations. As a guide these intervals may be regarded as warning or rejection for multiple 2SD outliers, or rejection for individual 3SD outliers in QC monitoring, although their precise application should be at the discretion of the QC manager concerned. A second method utilises a 5% window calculated directly from the certified value. Standard deviation is also shown in relative per cent for one, two and three relative standard deviations (1RSD, 2RSD and 3RSD) to facilitate an appreciation of the magnitude of these numbers and a comparison



with the 5% window. Caution should be exercised when concentration levels approach lower limits of detection of the analytical methods employed as performance gates calculated from standard deviations tend to be excessively wide whereas those determined by the 5% method are too narrow.

			Absolute	Standard	Deviation	3	Relative	Standard D	eviations	5% window	
Constituent	Certified Value	190	2SD	2SD	3SD	3SD	1000		2000		High
		130	Low	High	Low	High	IKOD	2830	3830	LOW	підп
Fire Assay											
Ag, ppm	284	16	253	316	237	332	5.59%	11.18%	16.76%	270	299
Au, ppm	5.18	0.151	4.87	5.48	4.72	5.63	2.91%	5.83%	8.74%	4.92	5.43
4-Acid Digest	tion	T	I	I	I	T	I	1	1	Γ	I
Ag, ppm	298	8	281	314	273	322	2.73%	5.47%	8.20%	283	312
AI, wt.%	3.98	0.156	3.67	4.29	3.51	4.45	3.91%	7.82%	11.73%	3.78	4.18
As, ppm	1801	119	1563	2039	1443	2158	6.61%	13.23%	19.84%	1711	1891
Be, ppm	0.71	0.08	0.55	0.87	0.47	0.95	11.38%	22.76%	34.14%	0.67	0.75
Bi, ppm	149	13	123	176	109	189	8.88%	17.76%	26.64%	142	157
Ca, wt.%	0.318	0.015	0.289	0.347	0.274	0.362	4.59%	9.18%	13.77%	0.302	0.334
Cd, ppm	54	3.3	47	60	44	63	6.09%	12.18%	18.26%	51	56
Ce, ppm	25.7	3.7	18.3	33.1	14.7	36.8	14.35%	28.69%	43.04%	24.4	27.0
Co, ppm	15.3	0.96	13.4	17.2	12.4	18.2	6.26%	12.53%	18.79%	14.5	16.1
Cr, ppm	30.2	5.3	19.7	40.8	14.4	46.1	17.47%	34.93%	52.40%	28.7	31.7
Cs, ppm	1.66	0.071	1.52	1.80	1.45	1.88	4.29%	8.57%	12.86%	1.58	1.75
Cu, wt.%	1.00	0.034	0.94	1.07	0.90	1.10	3.37%	6.75%	10.12%	0.95	1.05
Dy, ppm	1.21	0.072	1.07	1.36	1.00	1.43	5.94%	11.89%	17.83%	1.15	1.27
Er, ppm	0.59	0.024	0.55	0.64	0.52	0.67	4.05%	8.09%	12.14%	0.57	0.62
Eu, ppm	< 1	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Fe, wt.%	2.92	0.135	2.65	3.19	2.52	3.33	4.60%	9.21%	13.81%	2.78	3.07
Ga, ppm	22.2	1.96	18.3	26.1	16.3	28.1	8.81%	17.63%	26.44%	21.1	23.3
Gd, ppm	1.77	0.107	1.56	1.99	1.45	2.09	6.04%	12.08%	18.13%	1.68	1.86
Hf, ppm	2.53	0.167	2.19	2.86	2.02	3.03	6.63%	13.26%	19.90%	2.40	2.65
Ho, ppm	0.21	0.020	0.17	0.25	0.15	0.27	9.35%	18.71%	28.06%	0.20	0.22
In, ppm	11.2	0.96	9.3	13.1	8.3	14.1	8.61%	17.22%	25.83%	10.6	11.8
K, wt.%	0.623	0.019	0.584	0.662	0.565	0.681	3.11%	6.21%	9.32%	0.592	0.654
La, ppm	11.9	2.3	7.3	16.6	5.0	18.9	19.45%	38.90%	58.34%	11.3	12.5
Li, ppm	18.9	1.30	16.3	21.5	15.1	22.8	6.85%	13.70%	20.55%	18.0	19.9
Lu, ppm	0.099	0.012	0.074	0.124	0.061	0.136	12.62%	25.24%	37.86%	0.094	0.104

Table 4. Performance Gates for OREAS 603.

Note: intervals may appear asymmetric due to rounding.



			Absolute	Standard	Deviations	3	Relative	Standard D	eviations	5% window	
Constituent	Certified Value	1SD	2SD	2SD High	3SD	3SD High	1RSD	2RSD	3RSD	Low	High
4-Acid Digest	ion continue	ed	LOW	riigii	LOW	riigii					
Mg, ppm	828	45	738	919	692	965	5.48%	10.95%	16.43%	787	870
Mn, ppm	133	11	111	154	100	165	8.16%	16.32%	24.49%	126	139
Mo, ppm	6.05	0.516	5.02	7.08	4.51	7.60	8.52%	17.05%	25.57%	5.75	6.36
Na, wt.%	0.428	0.015	0.398	0.459	0.383	0.474	3.56%	7.13%	10.69%	0.407	0.450
Nb, ppm	7.09	0.74	5.62	8.57	4.88	9.31	10.42%	20.83%	31.25%	6.74	7.45
Nd, ppm	11.4	1.1	9.1	13.7	8.0	14.8	10.03%	20.06%	30.10%	10.8	12.0
Ni, ppm	112	5	102	121	98	126	4.25%	8.49%	12.74%	106	117
P, ppm	534	31	472	597	440	628	5.86%	11.72%	17.57%	507	561
Pb, ppm	1908	125	1658	2158	1534	2282	6.54%	13.08%	19.62%	1813	2003
Pr, ppm	3.23	0.190	2.85	3.61	2.65	3.80	5.90%	11.81%	17.71%	3.06	3.39
Rb, ppm	23.6	1.91	19.8	27.5	17.9	29.4	8.09%	16.18%	24.28%	22.4	24.8
S, wt.%	3.71	0.161	3.39	4.03	3.23	4.19	4.34%	8.67%	13.01%	3.52	3.89
Sb, ppm	205	18	170	240	152	258	8.61%	17.23%	25.84%	195	215
Sc, ppm	4.04	0.50	3.04	5.04	2.54	5.54	12.37%	24.73%	37.10%	3.84	4.24
Se, ppm	60	9	42	78	33	87	14.92%	29.85%	44.77%	57	63
Sm, ppm	2.14	0.134	1.87	2.40	1.73	2.54	6.27%	12.55%	18.82%	2.03	2.24
Sn, ppm	12.9	0.98	10.9	14.8	9.9	15.8	7.63%	15.25%	22.88%	12.2	13.5
Sr, ppm	459	77	306	613	229	689	16.71%	33.43%	50.14%	436	482
Ta, ppm	< 1	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Tb, ppm	0.23	0.023	0.19	0.28	0.16	0.30	9.87%	19.74%	29.61%	0.22	0.24
Te, ppm	56	7	42	71	35	78	12.66%	25.31%	37.97%	53	59
Th, ppm	5.90	0.85	4.20	7.60	3.35	8.45	14.42%	28.85%	43.27%	5.61	6.20
Ti, wt.%	0.191	0.018	0.156	0.226	0.138	0.244	9.25%	18.51%	27.76%	0.182	0.201
TI, ppm	4.18	0.174	3.83	4.53	3.66	4.71	4.17%	8.33%	12.50%	3.97	4.39
U, ppm	2.71	0.144	2.42	2.99	2.27	3.14	5.33%	10.67%	16.00%	2.57	2.84
V, ppm	32.0	2.22	27.5	36.4	25.3	38.6	6.94%	13.88%	20.82%	30.4	33.6
W, ppm	14.0	2.6	8.8	19.2	6.2	21.8	18.47%	36.94%	55.41%	13.3	14.7
Y, ppm	5.64	0.431	4.78	6.50	4.35	6.93	7.64%	15.27%	22.91%	5.36	5.92
Yb, ppm	0.61	0.09	0.44	0.78	0.35	0.87	14.24%	28.48%	42.71%	0.58	0.64
Zn, wt.%	0.920	0.031	0.857	0.982	0.826	1.013	3.39%	6.77%	10.16%	0.874	0.966
Zr, ppm	78	6.0	66	90	60	96	7.70%	15.39%	23.09%	74	82

Table 4 continued.

Note: intervals may appear asymmetric due to rounding.



Table 4 continued.											
Constituent	Certified		Absolute	Standard	Deviations	3	Relative	Standard D	eviations	5% w	indow
Constituent	Value	1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Aqua Regia D	igestion	1	1		1		L		L		
Ag, ppm	293	13	267	319	254	332	4.40%	8.80%	13.20%	278	308
Al, wt.%	0.580	0.092	0.396	0.765	0.303	0.857	15.91%	31.82%	47.74%	0.551	0.609
As, ppm	1766	91	1584	1949	1493	2040	5.16%	10.33%	15.49%	1678	1855
Au, ppm	5.08	0.179	4.72	5.43	4.54	5.61	3.53%	7.07%	10.60%	4.82	5.33
B, ppm	< 10	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Be, ppm	< 0.5	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Bi, ppm	151	14	123	179	109	193	9.25%	18.50%	27.75%	143	158
Ca, wt.%	0.237	0.017	0.202	0.271	0.185	0.289	7.27%	14.54%	21.81%	0.225	0.249
Cd, ppm	55	3.6	48	62	44	66	6.53%	13.06%	19.59%	52	58
Ce, ppm	13.0	2.4	8.1	17.8	5.7	20.2	18.73%	37.45%	56.18%	12.3	13.6
Co, ppm	15.2	0.79	13.6	16.7	12.8	17.5	5.19%	10.38%	15.57%	14.4	15.9
Cr, ppm	29.2	2.47	24.3	34.1	21.8	36.6	8.46%	16.92%	25.38%	27.7	30.7
Cs, ppm	0.74	0.12	0.50	0.97	0.39	1.09	15.76%	31.53%	47.29%	0.70	0.77
Cu, wt.%	1.01	0.026	0.96	1.06	0.93	1.08	2.56%	5.11%	7.67%	0.96	1.06
Dy, ppm	0.56	0.06	0.43	0.68	0.36	0.75	11.54%	23.07%	34.61%	0.53	0.58
Er, ppm	0.22	0.03	0.17	0.27	0.14	0.30	11.87%	23.74%	35.61%	0.21	0.23
Eu, ppm	0.25	0.05	0.15	0.35	0.10	0.40	19.92%	39.84%	59.76%	0.24	0.26
Fe, wt.%	2.85	0.155	2.54	3.16	2.38	3.31	5.44%	10.88%	16.32%	2.71	2.99
Ga, ppm	6.58	0.84	4.91	8.26	4.08	9.09	12.69%	25.38%	38.07%	6.26	6.91
Gd, ppm	0.86	0.12	0.63	1.10	0.51	1.21	13.54%	27.08%	40.61%	0.82	0.91
Hf, ppm	0.46	0.07	0.32	0.60	0.25	0.67	15.00%	30.00%	45.00%	0.44	0.48
Hg, ppm	< 4	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Ho, ppm	0.096	0.013	0.071	0.122	0.058	0.135	13.44%	26.88%	40.32%	0.092	0.101
In, ppm	11.7	1.15	9.4	14.0	8.2	15.1	9.83%	19.66%	29.50%	11.1	12.3
K, wt.%	0.093	0.015	0.062	0.124	0.047	0.139	16.44%	32.87%	49.31%	0.088	0.098
La, ppm	< 10	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Li, ppm	4.62	0.82	2.99	6.26	2.17	7.08	17.71%	35.42%	53.13%	4.39	4.85
Lu, ppb	26	5	17	36	12	41	18.27%	36.54%	54.81%	25	28
Mg, ppm	517	61	395	639	334	700	11.82%	23.64%	35.46%	491	543
Mn, ppm	130	6	117	142	110	149	4.99%	9.99%	14.98%	123	136
Mo, ppm	6.10	0.564	4.98	7.23	4.41	7.80	9.23%	18.47%	27.70%	5.80	6.41
Na, ppm	275	58	160	390	102	447	20.96%	41.91%	62.87%	261	288

Note: intervals may appear asymmetric due to rounding.



	Certified		Absolute Standard Deviations Relative Standard D							eviations 5% window	
Constituent	Value	1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
Aqua Regia D	igestion cor	ntinued							I		
Nb, ppm	< 0.5	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
Nd, ppm	5.90	0.71	4.47	7.33	3.75	8.04	12.12%	24.23%	36.35%	5.60	6.19
Ni, ppm	113	3	106	119	103	122	2.84%	5.67%	8.51%	107	118
P, ppm	172	27	117	227	90	255	15.94%	31.89%	47.83%	164	181
Pb, ppm	1691	36	1618	1763	1582	1799	2.15%	4.29%	6.44%	1606	1775
Pr, ppm	1.57	0.25	1.06	2.07	0.81	2.32	16.08%	32.15%	48.23%	1.49	1.64
Rb, ppm	4.89	0.88	3.13	6.65	2.25	7.54	17.98%	35.96%	53.94%	4.65	5.14
S, wt.%	3.46	0.185	3.09	3.83	2.91	4.01	5.34%	10.67%	16.01%	3.29	3.63
Sb, ppm	160	24	111	209	87	233	15.27%	30.53%	45.80%	152	168
Sc, ppm	1.06	0.16	0.74	1.39	0.57	1.55	15.39%	30.78%	46.16%	1.01	1.12
Se, ppm	59	8	43	75	35	84	13.82%	27.64%	41.46%	56	62
Sm, ppm	1.12	0.13	0.86	1.37	0.73	1.50	11.51%	23.02%	34.53%	1.06	1.17
Sn, ppm	12.0	1.11	9.8	14.2	8.7	15.3	9.25%	18.50%	27.76%	11.4	12.6
Sr, ppm	46.4	6.1	34.2	58.6	28.1	64.7	13.16%	26.32%	39.47%	44.1	48.7
Tb, ppm	0.11	0.02	0.08	0.15	0.06	0.17	17.02%	34.05%	51.07%	0.11	0.12
Te, ppm	57	2.8	52	63	49	66	4.97%	9.93%	14.90%	54	60
Th, ppm	2.30	0.29	1.72	2.88	1.43	3.17	12.59%	25.17%	37.76%	2.19	2.42
Ti, ppm	98	13	71	124	58	138	13.67%	27.34%	41.01%	93	103
TI, ppm	4.07	0.60	2.87	5.27	2.27	5.87	14.77%	29.53%	44.30%	3.86	4.27
U, ppm	0.89	0.10	0.69	1.08	0.59	1.18	11.00%	21.99%	32.99%	0.84	0.93
V, ppm	10.0	1.1	7.8	12.3	6.7	13.4	11.14%	22.28%	33.42%	9.5	10.5
W, ppm	3.88	0.78	2.32	5.45	1.54	6.23	20.15%	40.30%	60.45%	3.69	4.08
Y, ppm	2.28	0.35	1.58	2.98	1.23	3.33	15.33%	30.67%	46.00%	2.17	2.39
Yb, ppm	0.17	0.02	0.14	0.21	0.12	0.23	11.11%	22.23%	33.34%	0.17	0.18
Zn, wt.%	0.911	0.034	0.843	0.979	0.809	1.013	3.72%	7.44%	11.17%	0.865	0.956
Zr, ppm	16.0	1.6	12.8	19.3	11.2	20.9	10.10%	20.20%	30.30%	15.2	16.8
Infrared Com	bustion										
S, wt.%	3.97	0.161	3.65	4.29	3.49	4.45	4.06%	8.12%	12.19%	3.77	4.17

#### Table 4 continued.

Note: intervals may appear asymmetric due to rounding.

**Tolerance Limits** (ISO Guide 3207) were determined using an analysis of precision errors method and are considered a conservative estimate of true homogeneity. The meaning of tolerance limits may be illustrated for copper by 4-acid digestion, where 99% of the time (1- $\alpha$ =0.99) at least 95% of subsamples ( $\rho$ =0.95) will have concentrations lying between



0.99 and 1.02wt.%. Put more precisely, this means that if the same number of subsamples were taken and analysed in the same manner repeatedly, 99% of the tolerance intervals so constructed would cover at least 95% of the total population, and 1% of the tolerance intervals would cover less than 95% of the total population (ISO Guide 35).

For gold by fire assay and by aqua regia digestion, the tolerance limits have been determined by INAA using the reduced analytical subsample method which utilises the known relationship between standard deviation and analytical subsample weight (Ingamells and Switzer, 1973). In this approach the sample aliquot is substantially reduced to a point where most of the variability in replicate assays should be due to inhomogeneity of the reference material and measurement error becomes negligible. In this instance a subsample weight of 1g was employed and the 1RSD of 0.23% calculated at a 30g charge weight (1.72% at 1g weights) confirms the high level of gold homogeneity in OREAS 603.

Au by fire assay is reported by 27 laboratories and the charge weights range from 20-40g. The most common charge weight used in this round robin was 30g (19 labs) and tolerance intervals have been calculated at this sample weight. For Au by aqua regia digestion, tolerance limits have been calculated at a 25g sample weight (mode from the 25-50g sample weights used at 13 laboratories).

The gold homogeneity of OREAS 603 has also been evaluated in a **nested ANOVA** of the round robin program. Each of the twenty-eight round robin laboratories received six samples per CRM and these samples were made up of paired samples from three different, non-adjacent sampling intervals. The purpose of the ANOVA evaluation is to test that no statistically significant difference exists in the variance between-units to that of the variance within-units. This allows an assessment of homogeneity across the entire prepared batch of OREAS 600. The test was performed using the following parameters:

- Gold fire assay 162 samples (27 laboratories each providing analyses on 3 pairs of samples);
- Gold aqua regia digestion 78 samples (13 laboratories each providing analyses on 3 pairs of samples);
- Null Hypothesis, H<sub>0</sub>: Between-unit variance is no greater than within-unit variance (reject H<sub>0</sub> if *p*-value < 0.05);</li>
- Alternative Hypothesis,  $H_1$ : Between-unit variance is greater than within-unit variance.

*P*-values are a measure of probability where values less than 0.05 indicate a greater than 95% probability that the observed differences in within-unit and between-unit variances are real. The dataset was filtered for both individual and laboratory data set (batch) outliers prior to the calculation of the *p*-value. This process derived *p*-values of 0.95 for Au by fire assay and 0.21 for Au by aqua regia digestion. Both p-values are insignificant and the Null Hypothesis is retained. Additionally, none of the other 115 certified values showed significant *p*-values.

It is important to note that ANOVA is not an absolute measure of homogeneity. Rather, it establishes whether or not the analytes are distributed in a similar manner throughout the packaging run of OREAS 603 and whether the variance between two subsamples from the same unit is statistically distinguishable to the variance from two subsamples taken from any two separate units. A reference material therefore, can possess poor absolute homogeneity yet still pass a relative homogeneity test if the within-unit heterogeneity is large and similar across all units.



Based on the statistical analysis of the results of the inter-laboratory certification program it can be concluded that OREAS 603 is fit-for-purpose as a certified reference material (see 'Intended Use' below).

# PARTICIPATING LABORATORIES

- 1. Accurassay, Thunder Bay, Ontario, Canada
- 2. Acme (BV), Santiago, Chile
- 3. Actlabs, Ancaster, Ontario, Canada
- 4. AH Knight, Spartanburg, SC, USA
- 5. ALS, Johannesburg, South Africa
- 6. ALS, Lima, Peru
- 7. ALS, Reno, Nevada, USA
- 8. ALS, Townsville, QLD, Australia
- 9. ALS, Val-d'or, Quebec, Canada
- 10. ALS, Vancouver, BC, Canada
- 11. Bureau Veritas Geoanalytical, Adelaide, SA, Australia
- 12. Bureau Veritas Geoanalytical, Kalgoorlie, WA, Australia
- 13. Bureau Veritas Geoanalytical, Perth, WA, Australia
- 14. Bureau Veritas Kalassay, Kalgoorlie, WA, Australia
- 15. Inspectorate (BV), Lima, Peru
- 16. Inspectorate (BV), Sparks, Nevada, USA
- 17. Intertek Genalysis, Adelaide, SA, Australia
- 18. Intertek Genalysis, Perth, WA, Australia
- 19. Intertek Testing Services, Cupang, Muntinlupa, Philippines
- 20. Intertek Testing Services, Shunyi, Beijing, China
- 21. MINTEK Analytical Services, Randburg, South Africa
- 22. PT Geoservices Ltd, Cikarang, Jakarta Raya, Indonesia
- 23. SGS de Mexico, Durango, Mexico
- 24. SGS Geosol Laboratorios Ltda, Vespasiano, Minas Gerais, Brazil
- 25. SGS Lakefield Research Ltd, Lakefield, Ontario, Canada
- 26. SGS South Africa Pty Ltd, Booysens, Gauteng, South Africa
- 27. Shiva Analyticals Ltd, Bangalore North, Karnataka, India
- 28. SRL (Bureau Veritas), Perth, WA, Australia

## PREPARER AND SUPPLIER

Certified reference material OREAS 603 is prepared, certified and supplied by:



ORE Research & Exploration Pty Ltd	Tel:	+613-9729 0333
37A Hosie Street	Fax:	+613-9729 8338
Bayswater North VIC 3153	Web:	www.ore.com.au
AUSTRALIA	Email:	info@ore.com.au

It is available in unit sizes of 10 and 60g (single-use laminated foil pouches) and 1kg (plastic jars).



# **INTENDED USE**

OREAS 603 is intended for the following uses:

- for the monitoring of laboratory performance in the analysis of analytes reported in Table 1 in geological samples;
- for the verification of analytical methods for analytes reported in Table 1;
- for the calibration of instruments used in the determination of the concentration of analytes reported in Table 1.

# STABILITY AND STORAGE INSTRUCTIONS

OREAS 603 has been prepared from gold-silver-copper bearing ore from Evolution Mining's Mount Carlton Operation in Queensland, Australia and blended with argillic altered rhyodacite waste rock. It contains reactive sulphide (3.97% S) and has been packaged under a nitrogen environment (single use laminated foil pouches only). In its unopened state and under normal conditions of storage the CRM has a shelf life beyond ten years. Its stability will be monitored at regular intervals and purchasers notified if any changes are observed.

# INSTRUCTIONS FOR CORRECT USE

The certified values for OREAS 603 refer to the concentration level in its packaged state. It should not be dried prior to weighing and analysis. The certified values for gold by fire assay and aqua regia digestion are applicable to charge/sample weights ranging 20-50g.

# HANDLING INSTRUCTIONS

Fine powders pose a risk to eyes and lungs and therefore standard precautions such as the use of safety glasses and dust masks are advised.

# TRACEABILITY

The analytical samples were selected in a manner to represent the entire batch of prepared CRM. This 'representivity' was maintained in each submitted laboratory sample batch and ensures the user that the data is traceable from sample selection through to the analytical results that underlie the consensus values. Each analytical data set has been validated by its assayer through the inclusion of internal reference materials and QC checks during analysis. The laboratories were chosen on the basis of their competence (from past performance in inter-laboratory programs) for a particular analytical method, analyte or analyte suite, and sample matrix. Most of these laboratories have and maintain ISO 17025 accreditation. The certified values presented in this report are calculated from the means of accepted data following robust statistical treatment as detailed in this report.





# LEGAL NOTICE

Ore Research & Exploration Pty Ltd has prepared and statistically evaluated the property values of this reference material to the best of its ability. The Purchaser by receipt hereof releases and indemnifies Ore Research & Exploration Pty Ltd from and against all liability and costs arising from the use of this material and information.

## QMS ACCREDITED

ORE Pty Ltd is accredited to ISO 9001:2008 by Lloyd's Register Quality Assurance Ltd for its quality management system including development, manufacturing, certification and supply of CRMs.



# **CERTIFYING OFFICER**

Craig Hamlyn (B.Sc. Hons - Geology), Technical Manager - ORE P/L

# REFERENCES

Ingamells, C. O. and Switzer, P. (1973), Talanta 20, 547-568.

ISO Guide 30 (1992), Terms and definitions used in connection with reference materials.

ISO Guide 31 (2000), Reference materials – Contents of certificates and labels.

ISO Guide 3207 (1975), Statistical interpretation of data - Determination of a statistical tolerance interval.

ISO Guide 35 (2006), Certification of reference materials - General and statistical principals.

